

## **Bid Specification**

### **Madison County EMS Patient Lift Device**

#### **Bid Proposal Must Include Device/System For 13 Ambulances**

\*Currently our ambulances are utilizing Stryker Power Pro Stretchers with traditional “antler-type” floor mounting systems.

#### **Requirements for System:**

- Device must have capabilities to load and unload up to 700lbs into the ambulance with little or no assistance from EMS personnel.
- Device must conform to AS/NZS -4535 for dynamic crash testing
- Device must conform to BS EN-1789 clauses 4.5.9 for dynamic crash testing
- Device to be compliant to IEC 60601-1-2
- Device to be SAE J3027
- Device must provide a linear guide when loading and unloading the cot
- Device must provide manual back-up in the event of power failure
- Stretcher fastener shall have inductive charging capabilities for stretcher
- Bids should include all cost, including installation by either the manufacturer or any third-party vendor

#### **Included Options**

- In-Service DVD
- Operations Manual
- The bid shall include all freight and delivery charges
- Estimated date of delivery
- The bid shall include the cost of time and travel for an authorized representative to come and train staff.
- The bid must provide a full 7-year warranty on all parts, labor and travel for all 13 devices.

**\* All bids should provide clarification on any exemptions within the bid specification.**

**\* Bid proposals must be received by May 15, 2017 at 1600 hrs. EST.**

**\* Madison County EMS reserves the right to reject any and all bids.**